


METAL HAND STAMP "3/8", THIS SIDE, IN 50 (12.70) HIGH CHARACTERS NEAR CENTER OF PANEL.



SIDING PREPARATION
TYPICAL DETAIL
SCALE: 3/8"=1"

NOTE:
1- LEAD LINED PANELS SHALL BE CONSTRUCTED FROM SINGLE UNPIERCED SHEET, FREE OF OXIDE INCLUSIONS, FISSURES, VOIDS AND ALL OTHER DETRIMENTAL DEFECTS. MATERIAL SHALL BE LEAD GRADE "C" TO COMPLY WITH FEDERAL SPECIFICATION - QQL201. HUTCH SHALL MEET ALL APPLICABLE FEDERAL, STATE AND LOCAL RADIATION GUIDE LINES: NBS HANDBOOK 111 AND 114, ANSI N432 AND DOE ORDER 5400.11

NO	CHANGE / DESCRIPTION	AC	TS	DATE
02	CHANGE TO SIZE-D. MAKE ROOM FOR 2 UNISTRUTS	AC	TS	04/04
03	CHANGED OUTER STEEL FROM 112 TO 12 GA.	ACC	TS	04/04
04	UPDATE LENGTH DIM TO 91M BEFORE 1.22M	ACC	TS	04/04
01	UPDATE SECTION	TS	TS	04/04

ITEM	SYMBOL / QTY	DESCRIPTION	MATERIAL / SPEC	REV
PARTS LIST				
A1663604		ARGONNE NATIONAL LABORATORY		
ADVANCED PHOTON SOURCE		H0 TRANSOM 3/8" x .91M x 1.00M		
DESIGN BY	DATE	DATE	DATE	DATE
T. SANCHEZ	11/01/04	04/04/04	04/04/04	04/04/04
DESIGNED BY	BY	BY	BY	BY
T. SANCHEZ	T. KUZAY	T. KUZAY	T. KUZAY	T. KUZAY
SCALE	3/16:1	SHEET 114		
DATE	04/04/04	P41050903-820301-04		