



NOTE:
LEAD PANEL SHALL BE CONSTRUCTED FROM SINGLE UNPIERCED SHEET, FREE OF OXIDE INCLUSIONS, FISSURES, VOIDS AND ALL OTHER DETRIMENTAL DEFECTS. MATERIAL SHALL BE LEAD GRADE "C" TO COMPLY WITH FEDERAL SPECIFICATION-QLL201. HUTCH SHALL MEET ALL APPLICABLE FEDERAL, STATE AND LOCAL RADIATION GUIDE LINES: NBS HANDBOOK 111 AND 114, ANS N43.2 AND DOE ORDER 5480.11.

A1664002	4D	P41050903-820309-02	HO TRIM & MISC. 3/8"		AS
A1878402	3D	P41050903-820409-02	HO TRIM & MISC. 1/2"		RECD
	2		JAM NUT 3/4 HEX	SS	2
	1		HEX HD BOLT 3/4 x 1 3/4 LG.	SS	2
LOG NO.	REV	APP/DATE	DESCRIPTION	MATERIAL / SPEC	QTY

PARTS LIST

ARGONNE NATIONAL LABORATORY
ADVANCED PHOTON SOURCE

HO HUTCH DOOR PANEL MID.
3/8" x 2.50M x 1.02M

DATE: 10/21/14
DRAWN BY: R. KRAKORA
CHECKED BY: SANCHEZ/KRAKORA

DATE: 10/21/14
DATE: 10/21/14
DATE: 10/21/14

SCALE: 1/8"=1
SHEET 1 of 1
DWG. NO: P41050903-870308-00

DATE	CHANGE DESCRIPTION	BY	CHKD	DATE
	REV 1/18/14			