




METAL HAND STAMP 1/4", THIS SIDE IN 50 (127.70) HIGH CHARACTERS NEAR CENTER OF PANEL.



NOTE:  
1- LEAD LINED PANELS SHALL BE CONSTRUCTED FROM SINGLE UNPIERCED SHEET, FREE OF OXIDE INCLUSIONS, FISSURES, VOIDS AND ALL OTHER DETRIMENTAL DEFECTS. MATERIAL SHALL BE LEAD GRADE "C" TO COMPLY WITH FEDERAL SPECIFICATION - QQL201. HUTCH SHALL MEET ALL APPLICABLE FEDERAL, STATE AND LOCAL RADIATION GUIDE LINES: NBS HANDBOOK 111 AND 114, ANS N43.2 AND DOE ORDER 5480.11

ITEM	QUANTITY	UNITS	DESCRIPTION	MATERIAL / SPEC	REV
<b>PARTS LIST</b>					
<b>A2286101</b>					
<b>ARGONNE NATIONAL LABORATORY</b>					
<b>ADVANCED PHOTON SOURCE</b>					
H1 MBES ROOF #3 1/4' x 4.00M x .25M					
01	CHANGE TO SIZE-D, MAKE ROOM FOR 2 UNISTRUTS & A&BER RBE1		1/23/97		
02	CHANGE DESCRIPTION		07	04B	BAJC
REVISIONS					
DATE	BY	DESCRIPTION	DATE	BY	DESCRIPTION
6/5/95	J.S. PALANCAR	DESIGN BY	6/29/95	D. SHU	DESIGN CHECK
6/15/95	T. SANCHEZ	BY	7/3/95	T. KUZAY	PROJECT MGR
6/15/95	T. SANCHEZ	DESIGNED BY			
6/15/95	T. SANCHEZ	DESIGNED BY			
SCALE: 5/32=1 SHEET: 114					

01	CHANGE TO SIZE-D, MAKE ROOM FOR 2 UNISTRUTS & A&BER RBE1	1/23/97
02	CHANGE DESCRIPTION	07
REVISIONS		
DATE	BY	DESCRIPTION
1/23/97		
07		